































Symbol	Standard/Law Reference	Standard/Law Title	Symbol Title	Explanatory Text
1 	ISO 13225-1, Clause 5.1.1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Manufacturer	Indicates the medical device manufacturer.
	ISO 7000-3082	Graphical symbols for use on equipment		
2 	ISO 15223-1, Clause 5.1.3	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Date of Manufacture	Indicates the date when the medical device was manufactured.
	ISO 7000-2497	Graphical symbols for use on equipment		
3 	ISO 15223-1, Clause 5.1.6	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Catalogue or model number	Indicates the manufacturer's catalogue number so the device can be identified.
	ISO 7000-2493	Graphical symbols for use on equipment		
4 	ISO 15223-1, Clause 5.1.7	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Serial Number	Indicates the manufacturer's serial number so that a specific device can be identified.
	ISO 7000-2498	Graphical symbols for use on equipment		
5 	ISO 15223-1, Clause 5.1.5	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Batch/Lot Code	Indicates the manufacturer's batch/lot code so that the batch or lot can be identified.
	ISO 7000-2492	Graphical symbols for use on equipment		
6 	ISO 15223-1, Clause 5.4.4	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.
	IEC 60601-1, Table D.1, Symbol 10	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance		
	ISO 7000-0434	Graphical symbols for use on equipment		
7 	ISO 15223-1, Clause 5.3.7	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Temperature Limit	Indicates the temperature limits to which the medical device can be safely exposed.
	ISO 7000-0632	Graphical symbols for use on equipment		
8 	ISO 15223-1, Clause 5.3.8	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Humidity Limitation	Indicates the range of humidity to which the medical device can be safely exposed.
	ISO 7000-2620	Graphical symbols for use on equipment		
9 	ISO 15223-1, Clause 5.3.4	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Keep dry	Indicates a medical device needs to be protected from moisture.
	ISO 7000-0626	Graphical symbols for use on equipment		
10 	ISO 15223-1, Clause 5.3.1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Fragile, handle with care	Indicates a medical device that can be broken or damaged is not handled carefully.
	ISO 7000-0621	Graphical symbols for use on equipment		
11 	ISO 15223-1, Clause 5.2.7	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Non-sterile	Indicates a medical device that has not been subjected to a sterilization process.
	ISO 7000-2609	Graphical symbols for use on equipment		
12 	ISO 15223-1, Clause 5.4.3	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Consult instructions for use	Indicates user needs to consult instructions for use.
	ISO 7000-1641	Graphical symbols for use on equipment		
13 	IEC 60601-1, Table D.2, Symbol 10	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	Refer to instruction manual/booklet	Indicates user needs to consult instructions for use.

14 	ISO 15223-1, Clause 5.3.9	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Atmospheric pressure limits	Indicates the range of atmospheric pressure to which the medical device can be safely exposed.
	ISO 7000-2621	Graphical symbols for use on equipment		
15 	EN 50419	Marking of Electrical and Electronic Equipment in accordance with Article 11(2) of Directive 2002/96/EC (WEEE).	WEEE Wheeled Bin	This product contains electrical and electronic components that may contain materials which, if disposed with general waste, could be damaging to the environment. Residents of the European Union must follow specific disposal or recycling instructions for this product. Residents outside the European Union must dispose or recycle this product in accordance with local laws or regulations that apply.
16 	IEC 60601-1, Table D.1, Symbol 20	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	Type BF applied part	Identifies a type BF applied part complying with IEC 60601-1.
17	Directive 93/42/EEC Articles 4, 11, 12, 17 Annex 12	Council Directive 93/42/EEC of 14 June 1993 concerning medical devices		
	Directive 93/68/EEC	CE Marking	The requirements for accreditation and market surveillance relating to the marketing of products; Medical Device Directive.	Signifies European technical conformity.
18 	IEC 60417-5172 Section 7.2.6	Class II equipment		
19 	ISO 7000-2616	Graphical symbols for use on equipment	External cord connected	Indicates that device is connected to an external power source.
20 	ISO 7000-5008	Graphical symbols for use on equipment	OFF (power)	To indicate disconnection from the power.
21 	ISO 7000-5007	Graphical symbols for use on equipment	ON (power)	To indicate connection to power.
22 	ISO 7000-5417	Graphical symbols for use on equipment	Programmable duration	To identify the control of a programmable timer to start an operation at a specific point in time and to stop the operation at a specific point in time or after a specific duration; or to identify a display of the programmed or to-be-programmed duration
23 	ISO 7000-0623	Graphical symbols for use on equipment -- Registered symbols	This way up	To indicate correct upright position of the transport package.
24 	ISO 15223-1, Clause 5.1.9	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Distributor	Indicates the entity distributing the medical device into the locale. This symbol shall be accompanied by the name and address of the importing entity, adjacent to the symbol.
25 	ISO 15223-1, Clause 5.7.7	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Medical device	Indicates the item is a medical device.

26 	ISO 15223-1, Clause 5.7.10	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied	Unique Device Identifier	Indicates a carrier that contains Unique Device Identifier information. The use of this symbol is optional, but may be used when multiple data carriers are present on the label. If used, this symbol shall be placed adjacent to the Unique Device Identifier carrier. NOTE: Used to identify which information is associated with Unique Device Identifier
27 	IEC 60417-5172 Clause 5009	Graphical symbols for use on equipment	Stand-by	To identify the switch or switch position by means of which part of the equipment is switched on in order to bring it into the stand-by condition, and to identify the control to shift to or to indicate the state of low power consumption. Each of different states of power consumption may be indicated using a corresponding color
28 	IEC 60417: 2004 Clause 5016	Graphical symbols for use on equipment	Fuse	To identify fuse boxes or their location.
29 	IEC 60417: 2004 Clause 5032	Graphical symbols for use on equipment	Alternating current	To indicate on the rating plate that the equipment is suitable for alternating current only; to identify relevant terminals
30 	ISO 7000-1321B	Graphical symbols for use on equipment	Mass	To identify a function related to mass.